

EQUIPMENT

Laryngoscope Preventative Maintenance

Preventative Maintenance

- Clean/sterilize blades and ensure that there is no build-up or debris on blades or in areas of connection between blade and handle
- Scan lens for cracks
- Keep lens clean of debris using soft cloth and water or ethanol solution
- Check/replace bulb and battery as needed

