

Respiratory Monitor Preventative Maintenance

Preventative Maintenance

- Keep belt devoid of dirt and debris. Clean regularly with soap and water
- Clean electrodes after each use
- Check all cords for defects and replace as needed

Regularly check device-measured respiratory rate with a manual measurement