

EQUIPMENT

Otoscope Preventative Maintenance

Preventative Maintenance

- Ensure use of proper tips that are clean and free of defects
- Make sure tips are connected tightly for each use
- Regularly recharge or replace battery, as needed
- Clean lenses with soft cloth or cotton swab, and water or ethanol solution
- Remove debris from any rotating parts, use a thin, rigid member if necessary
- If using hand pump, ensure connections are tight and tubing is free of defects